

4

3


ONE NO. 4535 670 33181 SRT 1

1

PROCUREMENT	PART NUMBER
APPROVED SUPPLIER	110-0099-901
SPECTRA-STRIP	

NOTES: UNLESS OTHERWISE SPECIFIED

1. MATERIAL: 148-678 (RG-174A/U NON-MAGNETIC COAXIAL CABLE)
2. DIMENSIONS ARE IN INCHES.
3. ADD WHITE STRIPE TO ONE OUTER CONDUCTOR AS SHOWN.
4. CABLES TO BE SECURED TOGETHER WITH THE/WINYL ADHESIVE.
5. ALL TWELVE INDIVIDUAL CONDUCTORS TO BE CONFIGURED INTO INTO A FLAT RIBBON TYPE CABLE AS SHOWN.
6. PACKAGE IN 100 FT SPOOLS. LABEL SPOOLS WITH PHILIPS P/N 453567033181.



4

3

ONE NO. 4535 670 33181 SRT 1

1

REVISIONS					
REV.	ECO	DESCRIPTION	DRN BY	ENGR.	DATE
A	2005106	FIRST ISSUE	B.PETOT	B.PETOT	6/6/03
B	2005928	CORRECT VENDOR PART NUMBER	A.STEFANCHIK	D.MESSINE	09/19/03

DIGITAL SIGNATURES

Digitally signed by Dean Prud'homme  
DN: cn=Dean Prud'homme, ou=Philips Medical Systems, ou=Philips Medical Systems, c=US  
Date: 2003.09.30 11:53:20 -0400  
Reason: Document is released  
Not Valid

Digitally signed by Albert Stankovic  
DN: cn=Albert Stankovic, ou=Philips Medical Systems, ou=Philips Medical Systems, c=US  
Date: 2003.09.30 10:29:09 -0400  
Reason: I am the author of this document  
Not Valid

UNLESS OTHERWISE SPECIFIED - DIMENSIONS IN IN	
TOLERANCE	
X (+ N/A )	
XX (+ N/A )	
XXX (+ N/A )	
X (- N/A )	
XX (- N/A )	
XXX (- N/A )	
NAME: PHILIPS MEDICAL SYSTEMS (CLEVELAND), INC. 30800 CLEVELAND, OHIO 44143 USA	
CABLE, 12 CONDUCTOR RIBBON COAX	
DO NOT SCALE	
SCALE: 1:1	
TOOL: ACAD	
SRT: 1	QF: 1
REV: 1	VER: 1

2

1

A

B

C

D